

Mouse anti-PMS2

Cat. No.: AIB-30158 (0.5 ml Concentrate)

Instructions for use

Intended use

This antibody is designed for the specific localisation of PMS2 in formalin-fixed, paraffin-embedded tissue sections. Anti-PMS2 antibody is intended for research use only.

Specifications

| | |
|----------------------------|-------------------------------------|
| Specificity: | PMS2 |
| Immunogen: | Recombinant human PMS2 |
| Clone: | A16-4 |
| Isotype: | Mouse IgG1 kappa |
| Species reactivity: | Human +, mouse +, others not tested |

Summary and Description

PMS2 belongs to the mismatch repair proteins (MMR proteins) like MLH1, MLH3, MSH2, MSH3, MSH4, MSH5, MSH6, and PMS1. The PMS2 gene is homologue to the prokaryotic MutL gene.

Defects in the genes coding for MMR proteins lead to microsatellite instability (MSI) and considerable higher mutation rates. Hereditary nonpolyposis colorectal cancers (HNPCC) often show germline mutations in the mmr protein associated genes. These mutations result in decreased or abnormal protein production. The majority of HNPCC is characterised by damages in the MLH1 and MSH2 encoding genes, sometimes also in damages in MSH6 and PMS2 encoding genes.

Reagent provided

Mouse monoclonal antibody in buffer with carrier protein and preservative for stabilisation in the following formats:

Concentrate: 0.5 ml (Cat. No. AIB-30158)

Dilution of primary antibody

Dilution of Nordic Biosite' concentrated antibody depends on the detection system used. The final working dilution must always be determined by the user. The elaboration of staining protocol should be done by an experienced specialist. For Nordic Biosite' recommendations see chapter 'Staining procedure'.

Storage and handling

The antibody should be stored at 2-8°C without further dilution.

Dilutions of the concentrated antibody should be done with a suitable antibody dilution buffer (e.g. BCB-20005/BCB-20006 from Nordic Biosite). The diluted antibody should be stored at 2-8°C after use. Stability of this working solution depends on various parameters and has to be confirmed by appropriate controls. The antibody provided is suitable for use until the expiry date indicated on the label, if stored at 2-8°C. Do not use product after the expiry date.








Positive and negative controls should be run simultaneously with all specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact Nordic Biosite' technical support or your local distributor.

Precautions

Use through qualified personnel only.

Wear protective clothing to avoid contact of reagents and specimens with eye, skin and mucous membranes. If

Explanations of the symbols on the product label:

| | | | | |
|---|---|---|--|---|
|  | Catalog Number Bestellnummer Reference du catalogue |  | Batch Code Chargenbezeichnung Code du lot |  Manufacturer Nordic BioSite AB Propellervägen 4A S-183 62 Täby Sweden Tel: +46 (0)8 5444 33 40 Fax: +46 (0)8 756 94 90 info@nordicbiosite.com www.nordicbiosite.com |
|  | Use By Verwendbar bis Utiliser jusqu'à |  | In Vitro Diagnostic Medical Device In vitro Diagnostikum Dispositif médical de diagnostic in vitro | |
|  | Consult Instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation |  | Temperature Limitation Lagerungstemperatur Limites de température | |

reagents or specimens come in contact with sensitive area, wash with large amounts of water. Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur. ProClin300 is used for stabilisation. Material safety data sheets (MSDS) are available upon request.

Staining procedure

Refer to the following table for conditions specifically recommended for this antibody. Also refer to detection system data sheets for guidance on specific staining protocols or other requirements.

| Parameters | Nordic BioSites recommendations |
|---------------------------------|--|
| *Pre-treatment 20015/-20016) | Heat Induced Epitope Retrieval (for example in Citrate Buffer pH 6.0 (BCB- |
| *Control tissue | Colon mucosa |
| *Working dilution | 1:25-1:100 (for concentrates) |
| *Incubation time | 30-60 minutes |

Quality control

The recommended positive control tissue for this antibody is colon mucosa. We recommend carrying out a positive and a negative control with every staining run. Please refer to the instructions of the detection system for guidance on general quality control procedures.

Troubleshooting

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, refer to the instructions of the detection system for relevant information or contact your local distributor.

Expected results

This antibody stains positive in nuclei of PMS2 expressing cells in formalin-fixed, paraffin-embedded tissue sections. When staining colon carcinomas for PMS2 expression adjacent normal colon mucosa can be used as internal positive control. Interpretation of the staining results is solely the responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic procedure.

Limitations of the Procedure

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Endogenous peroxidase, alkaline phosphatase or biotin may cause non-specific staining depending on the detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive results with HRP (horse radish peroxidase) detection systems (Omata et al, 1980). Inadequate counterstaining and mounting can influence the interpretation of the results.

Nordic Biosite warrants that the product will meet all requirements described from its shipping date until the expiry date is reached, if the product is stored and utilised as recommended. No additional guarantees can be given. Under no circumstances shall Nordic Biosite be liable for any damages arising out of the use of the reagent provided.

Performance characteristics

Nordic Biosite has conducted studies to evaluate the performance of the antibody for use with a standard detection system. The product has been found to be sensitive and specific to the antigen of interest with minimal or no cross-reactivity.

Bibliography

Rüschhoff J et al. Pathologie 25:178-192, 2004
Umar A et al. J Natl Cancer Inst 96:261-268, 2004
Umar A et al. Natl Rev Cancer 4:153-158, 2004
Müller A et al. J Mol Diagnostics 6:308-315, 2004








Young J et al. Am J Pathol 159:2107-2116, 2001
Nadji M and Morales AR Ann N.Y. Acad Sci 420:134-9, 1983
Omata M et al. Am J Clin Pathol 73(5): 626-32, 1980

October 31, 2013

Rev: A1013

Doc: DBE_AIB-30158

Explanations of the symbols on the product label:

| | | | | |
|---|---|---|--|---|
|  REF | Catalog Number Bestellnummer Reference du catalogue |  LOT | Batch Code Chargenbezeichnung Code du lot |  Manufacturer Nordic BioSite AB Propellervägen 4A S-183 62 Täby Sweden Tel: +46 (0)8 5444 33 40 Fax: +46 (0)8 756 94 90 info@nordicbiosite.com www.nordicbiosite.com |
|  | Use By Verwendbar bis Utiliser jusque |  IVD | In Vitro Diagnostic Medical Device In vitro Diagnostikum Dispositif médical de diagnostic in vitro | |
|  | Consult Instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation |  | Temperature Limitation Lagerungstemperatur Limites de température | |