



## Mouse anti-Myosin SM Heavy Chain

Cat. No.: AIB-30202 (0.5 ml Concentrate);  
AIB-30201 (6 ml Ready-to-use)

### Instructions for use

#### Intended use

This antibody is designed for the specific localisation of the heavy chain of smooth muscle Myosin (Myosin SM) in formalin-fixed, paraffin-embedded tissue sections and in frozen sections.

Anti-Myosin SM Heavy Chain antibody is intended for in vitro diagnostic use.

#### Specifications

<b>Specificity:</b>	Myosin SM Heavy Chain
<b>Immunogen:</b>	SMMS-1
<b>Clone:</b>	col-1
<b>Isotype:</b>	Mouse IgG1 kappa
<b>Species reactivity:</b>	Human +, others not tested

#### Summary and Description

Smooth muscle myosin, heavy chain (SMMS) is a cytoplasmic structural protein that is a major component of the contractile apparatus of smooth muscle cells. Smooth muscle myosin is also a myoepithelium-associated protein. The antibody clone SMMS-1 reacts with human visceral and vascular smooth muscle cells and with myoepithelial cells. It is very helpful in distinguishing between benign sclerosing breast lesions and infiltrating breast carcinomas.

#### Reagent provided

Mouse monoclonal antibody from cell culture supernatant in PBS pH 7.4 with carrier protein and preservative for stabilisation in the following formats:

<b>Concentrate:</b>	0.5 ml	(Cat. No. AIB-30202)
<b>Ready-to-use:</b>	6 ml	(Cat. No. AIB-30201)

#### Dilution of primary antibody

Dilution of Nordic Biosite' concentrated antibody depends on the detection system used. The final working dilution must always be determined by the user. The elaboration of staining protocol should be done by an experienced specialist. For Nordic Biosite' recommendations see chapter 'Staining procedure'.








#### Storage and handling

The antibody should be stored at 2-8°C without further dilution.

Dilutions of the concentrated antibody should be done in a suitable antibody dilution buffer (e.g. BCB-20005/BCB-20006 from Nordic Biosite). The diluted antibody should be stored at 2-8°C after use. Stability of this working solution depends on various parameters and has to be confirmed by appropriate controls. The antibody provided is suitable for use until the expiry date indicated on the label, if stored at 2-8°C. Do not use product after the expiry date.

Positive and negative controls should be run simultaneously with all specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact Nordic Biosite' technical support or your local distributor.

Explanations of the symbols on the product label:

	Catalog Number Bestellnummer Reference du catalogue		Batch Code Chargenbezeichnung Code du lot	 Manufacturer  Nordic BioSite AB Propellervägen 4A S-183 62 Täby Sweden Tel: +46 (0)8 5444 33 40 Fax: +46 (0)8 756 94 90 info@nordicbiosite.com www.nordicbiosite.com
	Use By Verwendbar bis Utiliser jusqu'à		In Vitro Diagnostic Medical Device In vitro Diagnostikum Dispositif médical de diagnostic in vitro	
	Consult Instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation		Temperature Limitation Lagerungstemperatur Limites de température	

## Precautions

Use through qualified personnel only.

Wear protective clothing to avoid contact of reagents and specimens with eye, skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with large amounts of water.

Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur.

Sodium azide (NaN<sub>3</sub>), used for stabilisation, is not considered hazardous material in the concentration used.

Reaction of sodium azide with lead or copper in drainage pipes can result in the formation of highly explosive metallic azides. Sodium azide should be discarded in a large volume of running water to avoid formation of deposits. Material safety data sheets (MSDS) are available upon request.

## Staining procedure

Refer to the following table for conditions specifically recommended for this antibody. Also refer to detection system data sheets for guidance on specific staining protocols or other requirements.

### Parameters

\*Pre-treatment  
20015/-20016)  
\*Control tissue  
\*Working dilution  
\*Incubation time

### Nordic BioSites recommendations

Heat Induced Epitope Retrieval (for example in Citrate Buffer pH 6.0 (BCB-  
Breast tissue, intestinal tissue  
1:100-1:500 (for concentrates)  
30-60 minutes

## Quality control

The recommended positive control tissues for this antibody are breast or intestinal tissue. We recommend carrying out a positive and a negative control with every staining run. Please refer to the instructions of the detection system for guidance on general quality control procedures.

## Troubleshooting

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, refer to the instructions of the detection system for relevant information or contact your local distributor.

## Expected results

This antibody stains positive in the cytoplasm in formalin-fixed, paraffin-embedded tissue sections. Further details about the expression pattern of Myosin SM Heavy Chain can be found in the chapter 'Summary and Description'. Interpretation of the staining results is solely the responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic procedure.

## Limitations of the Procedure








Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Endogenous peroxidase, alkaline phosphatase or biotin may cause non-specific staining depending on the detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive results with HRP (horse radish peroxidase) detection systems (Omata et al, 1980). Inadequate counterstaining and mounting can influence the interpretation of the results.

Nordic Biosite warrants that the product will meet all requirements described from its shipping date until the expiry date is reached, if the product is stored and utilised as recommended. No additional guarantees can be given. Under no circumstances shall Nordic Biosite be liable for any damages arising out of the use of the reagent provided.

## Performance characteristics

Nordic Biosite has conducted studies to evaluate the performance of the antibody for use with a standard detection system. The product has been found to be sensitive and specific to the antigen of interest with minimal or no cross-reactivity.

Explanations of the symbols on the product label:

	Catalog Number Bestellnummer Reference du catalogue		Batch Code Chargenbezeichnung Code du lot	 Manufacturer  Nordic BioSite AB Propellervägen 4A S-183 62 Täby Sweden Tel: +46 (0)8 5444 33 40 Fax: +46 (0)8 756 94 90 info@nordicbiosite.com www.nordicbiosite.com
	Use By Verwendbar bis Utiliser jusque		In Vitro Diagnostic Medical Device In vitro Diagnostikum Dispositif médical de diagnostic in vitro	
	Consult Instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation		Temperature Limitation Lagerungstemperatur Limites de température	

## Bibliography

Werling RW et al. Am J Surg Pathol 27:82-90, 2003  
 Popnikolov NK et al. Am J Clin Pathol. 120:161-167, 2003  
 Agoff SN et al. Appl Immunohistochem Mol Morphol 9:164-169, 2001  
 Dabbs DJ, Gown AM. Diagn Cytopathol 20:203-207, 1999








Wang NP et al. Appl Immunohistochem 5:141-151, 1997  
 Lazard D et al. Proc Natl Acad Sci USA 90:999-1003, 1993  
 Nadjji M and Morales AR Ann N.Y. Acad Sci 420:134-9, 1983  
 Omata M et al. Am J Clin Pathol 73(5): 626-32, 1980

November 11, 2013

Rev: A1113

Doc: DBE\_AIB-30202/-30201

Explanations of the symbols on the product label:

	Catalog Number Bestellnummer Reference du catalogue		Batch Code Chargenbezeichnung Code du lot	 Manufacturer  Nordic BioSite AB Propellervägen 4A S-183 62 Täby Sweden Tel: +46 (0)8 5444 33 40 Fax: +46 (0)8 756 94 90 info@nordicbiosite.com www.nordicbiosite.com
	Use By Verwendbar bis Utiliser jusqu'à		In Vitro Diagnostic Medical Device In vitro Diagnostikum Dispositif médical de diagnostic in vitro	
	Consult Instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation		Temperature Limitation Lagerungstemperatur Limites de température	