

Rabbit Anti-Human CD16a Monoclonal Antibody (Clone SP175)

CATALOG #:

M4750 0.1 ml rabbit monoclonal antibody purified by protein

A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.

M4752 0.5 ml rabbit monoclonal

antibody purified by protein A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.


M4754 1.0 ml rabbit monoclonal


antibody purified by protein A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.

M4751 7.0 ml pre-diluted rabbit

monoclonal antibody purified by protein A/G in TBS/1% BSA buffer pH 7.6 with less than 0.1% sodium azide. y (didite of 170)





Flow cytometric analysis of rabbit anti-CD16a (SP175) antibody in Jurkat (green) compare to negative control of rabbit IgG (blue)

Tel: 1-925-474-8440

Toll Free: 1-800-787-6896 Fax: 1-925-474-8469

INTENDED USE: For Research Use Only. Not for use in diagnostic procedures

CLONE: SP175

IMMUNOGEN: Synthetic peptide derived from the C-terminus of human CD16a protein.

IG ISOTYPE: Rabbit IgG
EPITOPE: Not determined

MOLECULAR WEIGHT: 45 kDa

SPECIES REACTIVITY: Human (tested). (See www.springbio.com for information on species reactivity predicted by

sequence homology.)

DESCRIPTION: CD16 is a cluster of differentiation found on the surface of natural killer cells, neutrophils or

polymorphonuclear leukocytes (PMN), monocytes and macrophages. There are two CD16 isoforms, CD16a and CD16b. CD16a is expressed in monocytes, macrophages and NK cells and

CD16b is expressed in neutrophils.

APPLICATIONS: Immunohistochemistry (IHC) and Flow Cytometry

IHC PROCEDURE: Specimen Preparation: Formalin-fixed, paraffin-embedded tissues are suitable for use with this

primary antibody.

Deparaffinization: Deparaffinize slides using xylene or xylene alternative and graded alcohols. **Antibody Dilution:** If using the concentrate format of this product, dilute the antibody 1:100. The dilutions are estimates; actual results may differ because of variability in methods and protocols. **Antigen Retrieval:** Boil tissue section in EDTA buffer, pH 8.0 for 10 min followed by cooling at

room temperature for 20 min.

Primary Antibody Incubation: Incubate for 10 minutes at room temperature.

Slide Washing: Slides must be washed in between steps. Rinse slides with PBS/0.05% Tween. **Visualization:** Detect the antibody as instructed by the instructions provided with the visualization

system.

IHC POSITIVE CONTROL: Lung, Tonsil

FLOW CYTOMETRY: Recommended starting protocol: Dilute the antibody 1:100. Incubate for 30 minutes at 4°C.

The dilution is an estimate; actual results may differ because of variability in methods and protocols.

Optimal dilution and procedure should be determined by the end user.

FLOW CYTOMETRY

POSITIVE CONTROL: Jurkat Cell Line CELLULAR LOCALIZATION: Membrane

STORAGE & STABILITY:

Store at 2-8°C. Do not freeze. The user must validate any other storage conditions. When properly stored, the reagent is stable to the date indicated on the label. Do not use the reagent beyond the expiration date.

There are no definitive signs to indicate instability of this product; therefore, positive and negative controls should be tested simultaneously with unknown specimens.

If unexpected results are observed which cannot be explained by variations in laboratory procedures and a problem with the reagent is suspected, contact Technical Support at spring.tech@ventana.roche.com.

WARNINGS & PRECAUTIONS:

- Avoid contact of reagents with eyes and mucous membranes. If reagents come into contact with sensitive areas, wash with copious amounts of water.
- 2. This product is harmful if swallowed.
- 3. Consult local or state authorities with regard to recommended method of disposal.
- 4. Avoid microbial contamination of reagents.

Tel: 1-925-474-8440 Toll Free: 1-800-787-6896

Fax: 1-925-474-8469