Human seminoma stained

with anti-PLAP antibody



Rabbit Anti-Human PLAP Monoclonal Antibody (Clone SP15)

CATALOG #:

M3150 0.1 ml rabbit monoclonal antibody supplied as tissue culture supernatant in TBS/1% BSA buffer pH 7.5 with

less than 0.1% sodium azide.

M3152 0.5 ml rabbit monoclonal antibody supplied as tissue

culture supernatant in TBS/1% BSA buffer pH 7.5 with

less than 0.1% sodium azide.

1.0 ml rabbit monoclonal antibody supplied as tissue M3154

culture supernatant in TBS/1% BSA buffer pH 7.5 with less than 0.1% sodium azide.

M3151 7.0 ml pre-diluted rabbit monoclonal antibody supplied

as tissue culture supernatant in TBS/1% BSA buffer pH

7.6 with less than 0.1% sodium azide.

INTENDED USE: For Research Use Only. Not for use in diagnostic procedures.

CLONE:

IMMUNOGEN: Recombinant protein encoding human placental alkaline phosphatase.

IG ISOTYPE: Rabbit IqG EPITOPE: Not determined **MOLECULAR WEIGHT** 60-70kDa

SPECIES REACTIVITY: Human (tested). (See www.springbio.com for information on species reactivity predicted by

sequence homology.)

DESCRIPTION: This antibody reacts with a membrane-bound isoenzyme of placental alkaline phosphatase (PLAP)

occurring in the placenta during the 3rd trimester of gestation. It is expressed in testicular germ cell tumors. Unlike germ cell tumors, PLAP-positive somatic cell tumors uniformly express epithelial

membrane antigen (EMA).

APPLICATIONS: Immunohistochemistry (IHC)

IHC PROCEDURE: Specimen Preparation: Formalin-fixed, paraffin-embedded tissues are suitable for use with this

Deparaffinization: Deparaffinize slides using xylene or xylene alternative and graded alcohols. Antibody Dilution: If using the concentrate format of this product, dilute the antibody 1:100. The dilutions are estimates; actual results may differ because of variability in methods and protocols. Antigen Retrieval: Boil tissue section in 10mM citrate buffer, pH 6.0 for 10 min followed by

cooling at room temperature for 20 min.

Primary Antibody Incubation: Incubate for 30 minutes at room temperature.

Slide Washing: Slides must be washed in between steps. Rinse slides with PBS/0.05% Tween. Visualization: Detect the antibody as instructed by the instructions provided with the visualization

system.

POSITIVE CONTROL: Seminoma, placenta **CELLULAR LOCALIZATION:** Cytoplasm, membrane STORAGE & STABILITY

Store at 2-8°C. Do not freeze. The user must validate any other storage conditions. When properly stored, the reagent is stable to the date indicated on the label. Do not use the reagent beyond the

expiration date.

There are no definitive signs to indicate instability of this product; therefore, positive and negative

controls should be tested simultaneously with unknown specimens.

If unexpected results are observed which cannot be explained by variations in laboratory procedures and a problem with the reagent is suspected, contact Technical Support at

spring.tech@ventana.roche.com.

WARNINGS & PRECAUTIONS:

- Avoid contact of reagents with eyes and mucous membranes. If reagents come into contact with sensitive areas, wash with copious amounts of water.
- This product is harmful if swallowed.
- Consult local or state authorities with regard to recommended method of disposal.
- Avoid microbial contamination of reagents.

4300 Hacienda Drive Pleasanton, CA 94588 www.springbio.com

Tel: 1-925-474-8440 Toll Free: 1-800-787-6896 Fax: 1-925-474-8469